


TV2 Room Pressure Monitor Quick Start Guide

1. The Care and Use guide for your TV2 Room Pressure Monitor can be downloaded here:
www.e2di.com/usersguides.html

Please print out the Users Guide.

Although you should use the Users Guide to familiarize yourself with the TV2 installation and use, you can use the quick-start instructions below, if you are in a hurry.

Your TV2 room pressure monitor comes with free software to copy the logged data from your TV2 to your computer. To download the free TView software go to [the download page under support at www.e2di.com](#) and click on TView 6.3 link Room Pressure Monitor Software, copy the software to your desktop, unzip it and click on 'Setup' to install.




Temperature/RH Sensor



TV2 sensor sockets




Pressure Sensor



QUICKSTART


Your TV2 can be installed and operating in just a few minutes by following these simple steps.

1. Connect the USB cable to a PC or wall plug and the System Parameter display will appear.




2. Touch 'Up' or 'Down' to select C° or F° and touch 'Enter'.
3. Touch 'Up' to select the date format (MM/DD/YY, DD/MM/YY or YY/MM/DD) and then 'Enter'. The date format defaults to the US style (MM/DD/YY).
4. Set the day, month, year, hour and minute in the appropriate field by scrolling through the possible values with 'Up' or 'Down' after moving to the appropriate field with 'Enter'.
(Note: It is possible to move to another field with the 'Right' button but your entries will not be recorded).
5. Touch 'Escape'. You will be asked to verify that you want to erase all data. Use 'Right' to highlight YES and touch 'Enter'.
6. Touch 'Escape' to return to the previous menu.

The TV2 Room Pressure Monitor is now operational



Setting up sensors


1. Plug the pressure sensor into the socket next to the small green relay on the end of the TV2 display and the temp/RH sensor next to it. Position the pressure sensor so that the positive and negative clear plastic tube openings are in different rooms to see the pressure differential.
2. After a few seconds, the QuickCheck Monitor will recognize the pressure sensor and change the display to the Sensor Setup menu. Move the cursor to the Sample Data field by touching 'Right'.



3. This is how often you want the sensor to log a pressure reading. Scroll to any value between 1 min and 24 hrs with 'Up' or 'Down', touching 'Enter' to move to the minute and second field, or accept the default by touching 'Enter'. The default value is ten minutes (00:10:00). For pressure sensors we recommend you set it for 15 seconds (00:00:15).
4. The averaging is set to Medium which is fine for almost all applications.
5. Set the chart display (Y axis) with 'Up' and 'Down'. **This does not set the alarm.** These settings have no effect on the alarm, what data is collected, or what is stored. It simply sets the upper and lower point of the Y-axis of the data chart shown on the TV2.
6. Enter the Calibrate Sensor menu and press enter to auto calibrate the pressure sensor. This 'zeros' out the sensor to remove any position error. This may have to be repeated once the pressure sensor is placed in its location.

Setting the sensor alarms

5. Set Alarms, if needed, by highlighting the 'Set Alarm' button on the Sensor Setup screen with 'Left' or 'Right' and touching 'Enter'.



6. Enable the alarm by touching 'Up' and then 'Enter'.
7. Set the trigger-alarm-time to any value between 0:00 and 99:00. This controls how long the audible alarm and the relay will be activated.


Note: If you are using a dialer set it for at least 00:30 (30 seconds) but not longer than 06:00 (Six minutes).

8. Scroll in the upper alarm data limit and delay time and then the lower alarm limit and delay with 'Up' and 'Down' touching 'Enter' to move to each field in turn.

Note: The delay time is how long the condition must continue before the alarm occurs.


9. Touch 'Escape' until you return to the QuickCheck display.
10. Repeat these steps for each sensor.

You are done!



QuickCheck


The QuickCheck screen indicates the safe/unsafe condition of room pressure, temperature and humidity at all times. If all the numbers are green the room is safe, but if any of them appear red the clean room should not be used that condition is corrected.




Condition	If Green	If Red
Pressure	OK	Do not use room until positive pressure is restored
Temperature	OK	Adjust thermostat
Humidity	OK	Check humidity control
Alarm	Alarms set	Enable alarms (sensor setup menu)
Max/Min	OK	Condition was unsafe in the past – Reset by touching
Battery Icon	Charged	No action required – it will charge itself

If any condition (pressure, temperature or humidity) is red, it is outside the safe conditions and the room should not be used.

Safe limits are set in the alarm menu of each sensor.



If the Minimum pressure setting is red the pressure dropped below the safe conditions at some point in the past. As long as the current pressure is green it is ok. Check the logged data to see what happened



If the Alarm is disabled the quick check will be unable to display red conditions. It should be enabled and alarm conditions set in the sensor setup menu.

